



America

# CERTIFICATE

No. U8 10 03 63179 005

**Holder of Certificate:** **STORZ & BICKEL GMBH & CO. KG**

Rote Strasse 1  
78532 Tuttlingen  
GERMANY

**Production Facility(ies):** 63179

**Certification Mark:**



**Product:** **General Medical Devices  
Vaporizer system**

**Model(s):** **Volcano Medic**

<b>Parameters:</b>	Rated Input Voltage:	110-120V
	Rated Frequency:	50/60Hz
	Rated Input Power:	120VA
	Protection Class:	I

**Remarks:**  
The certificate is valid for USA and Canada.  
Other countries might have other requirements,  
which were not part of this certification.

**Tested according to:** UL 60601-1:2003  
CAN/CSA C22.2 No. 601.1-M90

The product was voluntarily tested according to the relevant safety requirements and mentioned properties. It can be marked with the certification mark shown above. The certification mark must not be altered in any way. See also notes overleaf.

**Test report no.:** 71332097

**Date,** 2010-03-24

Page 1 of 1

